

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001185  
16:52 21/07/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:42  
STERILIZING 3:47  
136.0 C 220kPa 3:47  
136.1 C 223kPa 4:17  
136.0 C 222kPa 4:47  
136.4 C 225kPa 5:17  
136.3 C 224kPa 5:47  
136.4 C 224kPa 6:17  
136.1 C 225kPa 6:47  
136.4 C 224kPa 7:17  
136.4 C 224kPa 7:17  
VENTING 7:17  
AIR DRYING 8:09  
CYCLE COMPLETE 15:26  
-----

Digital Signature #  
C260074D033419B8  
-----