
Midmark M11 - v1.0.0
Total Cycles: 5434

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2025 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.0 F 0.0
2:00 176.0 F 2.3
4:00 176.0 F 5.8
6:00 194.1 F 10.0
8:00 226.4 F 7.1
10:00 222.5 F 7.4
12:00 235.2 F 11.7
14:00 233.3 F 8.2
16:00 245.1 F 12.9
18:00 243.1 F 11.9
20:00 254.6 F 17.8
22:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.5 F 28.9
2:00 272.4 F 29.6
3:00 272.8 F 29.6
4:00 272.4 F 29.3

Min 271.0 F 28.4
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 227.3 F 0.0
10:00 224.1 F 0.0
15:00 228.0 F 0.0
20:00 226.9 F 0.0
25:00 230.6 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:03:29
CYCLE COMPLETE

03/24/2025 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____