

-----  
Midmark M11 - v1.0.5  
Total Cycles: 315

Name: RITTER2  
Sterilizer ID: V2313228  
08/27/2021 10:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 126.5 F 0.0  
2:00 127.7 F 0.0  
4:00 148.7 F 0.0  
6:00 164.7 F 0.1  
8:00 173.9 F 0.6  
10:00 193.5 F 1.3  
12:00 212.5 F 2.5  
14:00 220.5 F 4.5  
16:00 228.6 F 6.9  
18:00 237.0 F 9.9  
20:00 245.4 F 12.9  
22:00 242.8 F 11.8  
24:00 240.4 F 10.5  
26:00 239.6 F 9.7  
28:00 250.0 F 15.1  
30:00 260.1 F 20.8  
32:00 269.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.7 F 29.1  
2:00 272.3 F 28.5  
3:00 271.5 F 28.3  
4:00 272.4 F 29.0  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
5:00 181.8 F 0.0  
10:00 192.2 F 0.0  
15:00 189.2 F 0.0  
20:00 187.4 F 0.0  
25:00 188.8 F 0.0  
30:00 189.8 F 0.0  
  
Drying Complete

FILLING: 1:02  
HEATING: 33:05  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:11:29  
CYCLE COMPLETE

08/27/2021 11:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_