
Midmark M11 - v1.0.5
Total Cycles: 3917

Name: RITTER2
Sterilizer ID: V2313228
09/09/2024 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.0 F 0.0
2:00 181.4 F 2.3
4:00 180.8 F 6.1
6:00 197.7 F 10.8
8:00 217.1 F 5.0
10:00 228.0 F 9.9
12:00 233.2 F 9.1
14:00 241.7 F 12.9
16:00 241.7 F 11.8
18:00 254.2 F 18.3
20:00 265.5 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.2
1:00 272.3 F 30.0
2:00 271.8 F 29.3
3:00 271.8 F 29.8
4:00 272.2 F 29.7

Min 271.2 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 227.6 F 0.0
10:00 226.4 F 0.0
15:00 222.1 F 0.0
20:00 219.6 F 0.0
25:00 219.1 F 0.0
30:00 221.3 F 0.0
Drying Complete

FILLING: 1:29
HEATING: 21:40
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:00:46
CYCLE COMPLETE

09/09/2024 01:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____