
Midmark M11 - v1.0.5
Total Cycles: 1543

Name: RITTER2
Sterilizer ID: V2313228
09/27/2022 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.8 F 0.0
2:00 149.9 F 2.1
4:00 160.6 F 5.8
6:00 192.8 F 11.0
8:00 215.9 F 5.1
10:00 234.1 F 10.7
12:00 231.7 F 7.9
14:00 241.7 F 11.4
16:00 246.3 F 13.8
18:00 259.2 F 20.9
20:00 270.7 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.4
2:00 271.9 F 29.4
3:00 272.5 F 29.4
4:00 271.6 F 29.1

Min 271.5 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 233.2 F 0.0
10:00 236.5 F 0.0
15:00 229.7 F 0.0
20:00 232.2 F 0.0
25:00 234.2 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:44
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:55
CYCLE COMPLETE

09/27/2022 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____