
Midmark M11 - v1.0.5
Total Cycles: 2981

Name: RITTER2
Sterilizer ID: V2313228
10/05/2023 02:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.8 F 0.0
2:00 159.5 F 2.7
4:00 187.0 F 6.9
6:00 209.7 F 11.0
8:00 222.1 F 6.2
10:00 233.9 F 11.3
12:00 233.0 F 8.7
14:00 241.5 F 10.8
16:00 245.6 F 13.7
18:00 257.4 F 20.1
20:00 268.2 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.1 F 29.7
2:00 271.6 F 29.8
3:00 271.8 F 29.4
4:00 272.1 F 30.0

Min 271.3 F 29.2
Max 272.4 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 228.4 F 0.0
10:00 229.9 F 0.0
15:00 224.2 F 0.0
20:00 214.6 F 0.0
25:00 212.9 F 0.0
30:00 209.5 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 21:08
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

10/05/2023 03:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____