
Midmark M11 - v1.0.0
Total Cycles: 1357

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 191.1 F 3.3
4:00 196.1 F 7.3
6:00 223.1 F 12.6
8:00 228.4 F 7.3
10:00 240.6 F 12.6
12:00 238.2 F 10.1
14:00 238.4 F 8.3
16:00 250.0 F 15.2
18:00 261.9 F 22.0
20:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.0 F 28.4
2:00 272.4 F 28.9
3:00 271.4 F 28.3
4:00 272.1 F 28.7

Min 271.4 F 28.2
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 202.0 F 0.0
10:00 201.8 F 0.0
15:00 200.1 F 0.0
20:00 196.0 F 0.0
25:00 194.4 F 0.0
30:00 188.9 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

11/10/2021 05:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____