
Midmark M11 - v1.0.0
Total Cycles: 2845

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2023 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.7 F 0.0
2:00 202.8 F 4.7
4:00 209.1 F 10.5
6:00 223.1 F 5.8
8:00 237.3 F 11.6
10:00 237.1 F 9.4
12:00 237.4 F 8.7
14:00 251.5 F 15.7
16:00 264.5 F 23.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.7 F 28.6
2:00 271.4 F 27.8
3:00 272.8 F 28.8
4:00 272.1 F 28.2

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.2 F 0.0
5:00 217.8 F 0.0
10:00 219.6 F 0.0
15:00 221.6 F 0.0
20:00 219.2 F 0.0
25:00 217.4 F 0.0
30:00 215.1 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 17:49
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:22
CYCLE COMPLETE

02/01/2023 06:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____