
Midmark M11 - v1.0.5
Total Cycles: 3120

Name: RITTER2
Sterilizer ID: V2313228
11/13/2023 06:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.5 F 0.0
2:00 109.9 F 0.5
4:00 127.4 F 2.6
6:00 165.5 F 6.0
8:00 181.4 F 10.0
10:00 210.0 F 7.4
12:00 220.2 F 7.5
14:00 231.7 F 11.9
16:00 231.7 F 8.5
18:00 241.5 F 11.8
20:00 243.9 F 12.8
22:00 255.9 F 19.1
24:00 266.7 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.1
1:00 272.8 F 29.8
2:00 271.3 F 29.0
3:00 272.3 F 29.9
4:00 272.0 F 29.4

Min 271.1 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.7 F 0.0
5:00 227.9 F 0.0
10:00 197.4 F 0.0
15:00 193.5 F 0.0
20:00 189.6 F 0.0
25:00 188.1 F 0.0
30:00 189.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:33
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:08
CYCLE COMPLETE

11/13/2023 07:10 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____