
Midmark M11 - v1.0.5
Total Cycles: 1332

Name: RITTER2
Sterilizer ID: V2313228
07/21/2022 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.9 F 0.0
2:00 183.5 F 2.1
4:00 204.4 F 6.9
6:00 219.0 F 11.9
8:00 228.7 F 7.6
10:00 239.5 F 12.8
12:00 239.7 F 11.1
14:00 238.5 F 10.2
16:00 251.8 F 16.8
18:00 263.2 F 23.5
20:00 272.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.7 F 29.7
2:00 271.9 F 29.0
3:00 271.9 F 29.5
4:00 272.4 F 29.7

Min 271.4 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.5 F 0.0
5:00 227.3 F 0.0
10:00 224.3 F 0.0
15:00 216.3 F 0.0
20:00 202.2 F 0.0
25:00 192.3 F 0.0
30:00 191.7 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

07/21/2022 01:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____