

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2764

Name: RITTER 1  
Sterilizer ID: V2178198  
01/11/2023 10:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
2:00 198.7 F 0.0  
4:00 210.2 F 5.3  
6:00 226.5 F 10.7  
8:00 226.4 F 6.5  
10:00 239.5 F 12.5  
12:00 239.4 F 10.5  
14:00 237.6 F 9.1  
16:00 252.6 F 16.4  
18:00 264.8 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.5 F 28.7  
2:00 271.5 F 28.2  
3:00 272.6 F 28.7  
4:00 271.6 F 28.0  
  
Min 271.5 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
5:00 226.2 F 0.0  
10:00 233.9 F 0.0  
15:00 231.0 F 0.0  
20:00 227.2 F 0.0  
25:00 226.6 F 0.0  
30:00 227.3 F 0.0  
  
Drying Complete

FILLING: 1:31  
HEATING: 19:50  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:32  
CYCLE COMPLETE

01/11/2023 11:39 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_