

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3858

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 166.7 F 0.0  
2:00 165.5 F 0.6  
4:00 178.6 F 3.8  
6:00 201.2 F 8.6  
8:00 218.7 F 13.5  
10:00 230.1 F 7.7  
12:00 236.8 F 12.5  
14:00 237.4 F 9.4  
16:00 238.9 F 9.3  
18:00 247.7 F 13.8  
20:00 259.2 F 20.1  
22:00 269.8 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 271.8 F 28.3  
2:00 272.4 F 28.8  
3:00 271.2 F 28.1  
4:00 271.8 F 28.4

Min 271.2 F 27.9  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 237.0 F 0.0  
10:00 219.0 F 0.0  
15:00 210.1 F 0.0  
20:00 186.1 F 0.0  
25:00 177.6 F 0.0  
30:00 199.2 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 22:55  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:02:03  
CYCLE COMPLETE

10/04/2023 10:05 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_