
Midmark M11 - v1.0.5
Total Cycles: 444

Name: RITTER2
Sterilizer ID: V2313228
10/13/2021 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.6 F 0.0
2:00 174.2 F 2.1
4:00 183.4 F 6.4
6:00 205.7 F 11.4
8:00 227.1 F 6.6
10:00 239.9 F 11.7
12:00 237.0 F 9.4
14:00 238.4 F 9.8
16:00 248.2 F 14.6
18:00 259.7 F 21.0
20:00 270.1 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.3 F 29.3
2:00 271.3 F 29.0
3:00 272.3 F 29.2
4:00 271.2 F 28.8

Min 271.2 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.1 F 0.0
5:00 236.0 F 0.0
10:00 221.5 F 0.0
15:00 225.3 F 0.0
20:00 221.1 F 0.0
25:00 216.5 F 0.0
30:00 214.2 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

10/13/2021 11:39 AM
ACCEPTED: X
REJECTED:
Operator: U01