
Midmark M11 - v1.0.0
Total Cycles: 832

Name: RITTER 1
Sterilizer ID: V2178198
05/20/2021 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 198.5 F 2.6
4:00 192.7 F 7.4
6:00 230.4 F 12.5
8:00 221.5 F 7.2
10:00 234.9 F 12.7
12:00 237.9 F 10.5
14:00 237.5 F 9.0
16:00 252.0 F 15.9
18:00 263.9 F 22.8
20:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.4
1:00 272.0 F 28.6
2:00 272.1 F 28.8
3:00 272.0 F 28.8
4:00 271.7 F 28.6

Min 271.6 F 28.4
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 242.3 F 0.0
10:00 230.2 F 0.0
15:00 226.6 F 0.0
20:00 233.0 F 0.0
25:00 228.1 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 2:05
HEATING: 20:35
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

05/20/2021 02:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____