
Midmark M11 - v1.0.5
Total Cycles: 2992

Name: RITTER2
Sterilizer ID: V2313228
10/10/2023 05:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.9 F 0.0
2:00 181.7 F 3.8
4:00 204.9 F 9.5
6:00 223.8 F 6.5
8:00 235.3 F 10.9
10:00 235.2 F 9.0
12:00 238.0 F 9.5
14:00 250.3 F 15.8
16:00 263.1 F 23.2
18:00 271.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.7 F 29.4
2:00 271.8 F 29.2
3:00 272.6 F 29.3
4:00 271.7 F 29.2

Min 271.5 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 235.7 F 0.0
10:00 223.7 F 0.0
15:00 217.5 F 0.0
20:00 208.3 F 0.0
25:00 200.4 F 0.0
30:00 201.8 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

10/10/2023 06:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____