
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000135
11:55 12/03/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:39
STERILIZING 4:38
136.0 C 220kPa 4:38
136.4 C 221kPa 5:08
136.2 C 221kPa 5:38
136.4 C 224kPa 6:08
136.5 C 224kPa 6:38
136.4 C 224kPa 7:08
136.4 C 223kPa 7:38
136.3 C 223kPa 8:08
136.3 C 223kPa 8:08
VENTING 8:08
AIR DRYING 9:01
CYCLE COMPLETE 9:07

Digital Signature #
05CFE6EBE69BAC5C
