
Midmark M11 - v1.0.0
Total Cycles: 2926

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2023 08:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.1 F 0.0
2:00 77.7 F 0.0
4:00 87.7 F 0.0
6:00 114.7 F 1.5
8:00 141.2 F 4.1
10:00 168.8 F 7.3
12:00 189.5 F 11.2
14:00 217.5 F 4.4
16:00 225.1 F 7.2
18:00 230.3 F 10.9
20:00 230.1 F 6.8
22:00 240.7 F 11.5
24:00 236.2 F 8.7
26:00 249.6 F 14.9
28:00 260.8 F 21.2
30:00 271.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 28.5
2:00 272.5 F 28.9
3:00 271.6 F 28.8
4:00 272.0 F 28.5

Min 271.3 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 224.4 F 0.0
10:00 191.0 F 0.0
15:00 181.6 F 0.0
20:00 181.9 F 0.0
25:00 182.5 F 0.0
30:00 182.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:19
CYCLE COMPLETE

02/22/2023 09:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____