

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1768

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2022 10:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.7 F 0.0  
2:00 170.5 F 2.8  
4:00 182.6 F 6.6  
6:00 211.4 F 12.1  
8:00 222.2 F 6.8  
10:00 235.9 F 12.0  
12:00 234.8 F 9.1  
14:00 239.2 F 9.4  
16:00 247.2 F 13.9  
18:00 259.5 F 20.7  
20:00 270.7 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.5  
1:00 272.6 F 29.2  
2:00 271.8 F 28.5  
3:00 271.8 F 29.0  
4:00 272.6 F 29.1  
  
Min 271.2 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.1 F 0.0  
5:00 224.3 F 0.0  
10:00 235.5 F 0.0  
15:00 231.8 F 0.0  
20:00 227.7 F 0.0  
25:00 230.2 F 0.0  
30:00 229.3 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:50  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

03/30/2022 11:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_