
Midmark M11 - v1.0.0
Total Cycles: 3009

Name: RITTER 1
Sterilizer ID: V2178198
03/15/2023 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.1 F 0.0
2:00 181.7 F 3.9
4:00 197.6 F 8.5
6:00 222.3 F 10.6
8:00 227.5 F 8.7
10:00 239.2 F 11.2
12:00 241.2 F 11.7
14:00 239.9 F 10.4
16:00 253.9 F 17.4
18:00 265.7 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.2
2:00 271.9 F 28.5
3:00 271.8 F 28.8
4:00 272.5 F 29.1

Min 271.5 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 217.5 F 0.0
10:00 224.6 F 0.0
15:00 229.5 F 0.0
20:00 229.7 F 0.0
25:00 228.3 F 0.0
30:00 226.7 F 0.0

Drying Complete

FILLING: 1:23
HEATING: 19:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:23
CYCLE COMPLETE

03/15/2023 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____