
Midmark M11 - v1.0.5
Total Cycles: 2428

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.8 F 0.0
2:00 190.4 F 4.3
4:00 207.4 F 10.0
6:00 220.0 F 6.5
8:00 234.0 F 12.0
10:00 236.8 F 10.0
12:00 236.5 F 8.7
14:00 251.1 F 16.1
16:00 263.0 F 23.2
18:00 272.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.6 F 29.8
2:00 271.8 F 29.0
3:00 272.4 F 29.9
4:00 272.4 F 29.6

Min 271.4 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 226.7 F 0.0
10:00 235.1 F 0.0
15:00 241.4 F 0.0
20:00 232.9 F 0.0
25:00 229.8 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 2:32
HEATING: 18:16
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

05/29/2023 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____