

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005981  
9:38 17/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:58  
PRESSURIZING 3:00  
STERILIZING 4:03  
136.5 C 223kPa 4:03  
136.1 C 222kPa 4:33  
136.2 C 224kPa 5:03  
136.1 C 221kPa 5:33  
136.4 C 224kPa 6:03  
135.9 C 221kPa 6:33  
136.4 C 224kPa 7:03  
136.2 C 224kPa 7:33  
136.2 C 224kPa 7:33  
VENTING 7:33  
AIR DRYING 8:26  
CYCLE COMPLETE 21:08

-----  
Digital Signature #  
7DA477E5C0FF6A37  
-----