
Midmark M11 - v1.0.0
Total Cycles: 2958

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2023 06:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.8 F 0.0
2:00 184.2 F 3.9
4:00 203.4 F 8.7
6:00 221.0 F 7.8
8:00 230.9 F 9.8
10:00 234.0 F 8.3
12:00 244.6 F 13.5
14:00 244.2 F 12.5
16:00 257.5 F 19.6
18:00 269.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.5 F 29.1
2:00 271.6 F 28.6
3:00 272.6 F 29.1
4:00 271.7 F 28.4

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 218.1 F 0.0
10:00 224.1 F 0.0
15:00 226.2 F 0.0
20:00 237.1 F 0.0
25:00 223.6 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:01
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

02/28/2023 07:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____