
Midmark M11 - v1.0.0
Total Cycles: 614

Name: MIDMARK
Sterilizer ID: V2178198
01/20/2021 02:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.6 F 0.0
2:00 202.1 F 3.7
4:00 214.8 F 10.4
6:00 225.2 F 6.7
8:00 242.0 F 13.5
10:00 241.8 F 11.8
12:00 241.0 F 10.7
14:00 255.5 F 18.2
16:00 267.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 271.8 F 28.2
2:00 271.9 F 28.3
3:00 272.4 F 28.6
4:00 272.6 F 28.9

Min 271.5 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 231.9 F 0.0
10:00 236.5 F 0.0
15:00 226.5 F 0.0
20:00 220.5 F 0.0
25:00 220.4 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 17:08
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 00:56:16
CYCLE COMPLETE

01/20/2021 03:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____