
Midmark M11 - v1.0.0
Total Cycles: 902

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2021 04:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.0 F 0.0
2:00 112.4 F 0.0
4:00 123.5 F 0.1
6:00 152.5 F 1.5
8:00 174.0 F 4.0
10:00 192.3 F 6.8
12:00 213.2 F 10.5
14:00 220.1 F 6.5
16:00 226.4 F 9.0
18:00 236.9 F 13.5
20:00 237.7 F 10.6
22:00 236.1 F 8.6
24:00 248.0 F 14.4
26:00 259.1 F 20.5
28:00 269.3 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.2
2:00 271.9 F 28.6
3:00 272.0 F 29.0
4:00 272.5 F 29.1

Min 271.5 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 241.3 F 0.0
10:00 226.2 F 0.0
15:00 224.4 F 0.0
20:00 223.3 F 0.0
25:00 220.5 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:07
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:35
CYCLE COMPLETE

06/15/2021 06:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____