
Midmark M11 - v1.0.0
Total Cycles: 2769

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 192.5 F 0.8
4:00 210.6 F 6.0
6:00 222.0 F 11.5
8:00 228.6 F 7.6
10:00 242.1 F 13.4
12:00 242.4 F 11.7
14:00 241.1 F 10.6
16:00 255.1 F 17.8
18:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.4 F 28.8
2:00 271.7 F 28.0
3:00 272.5 F 28.8
4:00 272.0 F 28.2

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 228.1 F 0.0
10:00 239.3 F 0.0
15:00 238.0 F 0.0
20:00 231.4 F 0.0
25:00 225.5 F 0.0
30:00 221.7 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

01/11/2023 05:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____