

-----  
Midmark M11 - v1.0.5  
Total Cycles: 357

Name: RITTER2  
Sterilizer ID: V2313228  
09/13/2021 01:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.5 F 0.0  
2:00 121.1 F 1.2  
4:00 135.4 F 3.6  
6:00 162.9 F 7.0  
8:00 187.0 F 10.8  
10:00 205.2 F 4.4  
12:00 220.7 F 7.0  
14:00 234.1 F 11.6  
16:00 235.1 F 8.9  
18:00 240.7 F 10.9  
20:00 246.6 F 13.8  
22:00 258.0 F 20.0  
24:00 268.5 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.8  
1:00 272.1 F 28.9  
2:00 272.4 F 29.4  
3:00 271.4 F 28.8  
4:00 272.2 F 29.2

Min 271.3 F 28.6  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.3 F 0.0  
5:00 230.0 F 0.0  
10:00 217.1 F 0.0  
15:00 212.1 F 0.0  
20:00 206.8 F 0.0  
25:00 211.9 F 0.0  
30:00 199.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 25:09  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:03:26  
CYCLE COMPLETE

09/14/2021 06:13 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01