

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 003262  
15:14 02/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:36  
STERILIZING 3:39  
136.4 C 220kPa 3:39  
136.4 C 224kPa 4:09  
136.2 C 222kPa 4:39  
136.4 C 225kPa 5:09  
136.3 C 224kPa 5:39  
136.3 C 224kPa 6:09  
136.3 C 223kPa 6:39  
136.4 C 222kPa 7:09  
136.4 C 222kPa 7:09  
VENTING 7:09  
AIR DRYING 8:03  
CYCLE COMPLETE 8:08

-----  
Digital Signature #  
0DBCEC6E5FF210CA  
-----