
Midmark M11 - v1.0.5
Total Cycles: 1703

Name: RITTER2
Sterilizer ID: V2313228
11/09/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.0 F 0.0
2:00 177.9 F 0.2
4:00 180.7 F 1.2
6:00 206.5 F 6.1
8:00 226.4 F 11.6
10:00 230.4 F 8.2
12:00 241.6 F 11.3
14:00 244.1 F 12.9
16:00 244.3 F 12.5
18:00 257.9 F 19.8
20:00 269.9 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.7 F 29.3
2:00 271.6 F 28.5
3:00 272.8 F 29.4
4:00 272.2 F 28.8

Min 271.4 F 28.5
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 219.1 F 0.0
10:00 229.4 F 0.0
15:00 230.1 F 0.0
20:00 220.5 F 0.0
25:00 215.4 F 0.0
30:00 214.5 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 20:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

11/09/2022 10:42 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____