
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007887
7:33 26/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:52
STERILIZING 4:01
136.4 C 220kPa 4:01
136.2 C 222kPa 4:31
136.0 C 220kPa 5:01
136.4 C 222kPa 5:31
136.0 C 221kPa 6:01
136.1 C 222kPa 6:31
136.3 C 220kPa 7:01
136.1 C 222kPa 7:31
136.1 C 222kPa 7:31
VENTING 7:31
AIR DRYING 8:21
CYCLE COMPLETE 36:47

Digital Signature #
77BECEC4D0FD4D87
