
Midmark M11 - v1.0.0
Total Cycles: 5387

Name: RITTER 1
Sterilizer ID: V2178198
03/05/2025 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.3 F 0.0
2:00 175.5 F 2.3
4:00 180.2 F 5.8
6:00 187.0 F 9.7
8:00 225.0 F 7.7
10:00 218.2 F 6.9
12:00 228.5 F 11.2
14:00 230.8 F 7.5
16:00 243.1 F 12.8
18:00 240.5 F 11.0
20:00 251.7 F 16.7
22:00 261.8 F 22.7
24:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 271.7 F 29.5
2:00 272.3 F 29.9
3:00 272.8 F 29.8
4:00 272.8 F 29.6

Min 271.1 F 28.9
Max 273.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 210.1 F 0.0
10:00 207.1 F 0.0
15:00 212.4 F 0.0
20:00 212.6 F 0.0
25:00 212.9 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

03/05/2025 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____