
Midmark M11 - v1.0.5
Total Cycles: 3048

Name: RITTER2
Sterilizer ID: V2313228
10/24/2023 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 134.9 F 0.0
2:00 136.4 F 1.3
4:00 154.2 F 3.9
6:00 178.4 F 8.0
8:00 202.7 F 13.0
10:00 217.4 F 7.2
12:00 235.3 F 12.1
14:00 232.9 F 8.5
16:00 240.3 F 11.2
18:00 244.1 F 12.7
20:00 256.6 F 19.3
22:00 267.5 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.1 F 29.1
2:00 272.3 F 29.6
3:00 271.1 F 28.9
4:00 272.4 F 29.3

Min 271.0 F 28.8
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 222.2 F 0.0
10:00 188.2 F 0.0
15:00 186.3 F 0.0
20:00 187.6 F 0.0
25:00 194.8 F 0.0
30:00 197.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:32
CYCLE COMPLETE

10/24/2023 10:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____