

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 002520  
15:57 19/02/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:19  
STERILIZING 3:18  
136.5 C 220kPa 3:18  
136.3 C 226kPa 3:48  
136.4 C 226kPa 4:18  
136.4 C 225kPa 4:48  
136.2 C 225kPa 5:18  
136.2 C 224kPa 5:48  
136.2 C 223kPa 6:18  
136.2 C 225kPa 6:48  
136.2 C 225kPa 6:48  
VENTING 6:48  
AIR DRYING 7:39  
CYCLE COMPLETE 7:45  
-----

Digital Signature #  
BBCCA51C90A3FC0C  
-----