
Midmark M11 - v1.0.0
Total Cycles: 1522

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.6 F 0.0
2:00 195.3 F 4.1
4:00 205.4 F 9.3
6:00 228.1 F 7.7
8:00 234.8 F 10.0
10:00 234.0 F 8.2
12:00 245.0 F 13.2
14:00 243.5 F 11.9
16:00 256.5 F 18.7
18:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.4 F 29.0
2:00 272.0 F 28.4
3:00 271.9 F 28.7
4:00 272.4 F 28.7

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 240.1 F 0.0
10:00 222.0 F 0.0
15:00 224.8 F 0.0
20:00 223.7 F 0.0
25:00 198.9 F 0.0
30:00 205.6 F 0.0

Drying Complete

FILLING: 2:15
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

01/13/2022 11:49 AM
ACCEPTED: X
REJECTED:
Operator: U02