

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001985  
9:10 03/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:46  
PRESSURIZING 0:55  
STERILIZING 2:09  
136.0 C 220kPa 2:09  
136.4 C 223kPa 2:39  
136.4 C 223kPa 3:09  
136.4 C 221kPa 3:39  
136.2 C 222kPa 4:09  
136.3 C 223kPa 4:39  
136.4 C 224kPa 5:09  
136.3 C 222kPa 5:39  
136.3 C 222kPa 5:39  
VENTING 5:39  
AIR DRYING 6:39  
CYCLE COMPLETE 6:43  
-----

Digital Signature #  
34D722F72516F407  
-----