
Midmark M11 - v1.0.5
Total Cycles: 1533

Name: RITTER2
Sterilizer ID: V2313228
09/26/2022 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 163.4 F 2.3
4:00 184.8 F 5.9
6:00 202.6 F 10.6
8:00 218.8 F 4.7
10:00 224.5 F 10.4
12:00 229.8 F 7.5
14:00 242.0 F 13.4
16:00 242.4 F 12.3
18:00 255.2 F 19.0
20:00 266.6 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.4 F 30.1
2:00 271.5 F 29.3
3:00 272.2 F 30.3
4:00 272.2 F 29.8

Min 271.4 F 29.3
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 203.1 F 0.0
10:00 220.1 F 0.0
15:00 217.8 F 0.0
20:00 227.8 F 0.0
25:00 227.1 F 0.0
30:00 233.0 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 21:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

09/26/2022 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____