
Midmark M11 - v1.0.0
Total Cycles: 3739

Name: RITTER 1
Sterilizer ID: V2178198
09/08/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 160.6 F 0.8
4:00 184.7 F 4.8
6:00 210.2 F 9.9
8:00 221.2 F 6.0
10:00 230.9 F 10.4
12:00 232.9 F 7.5
14:00 245.0 F 12.7
16:00 244.4 F 12.5
18:00 257.1 F 19.2
20:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.2 F 28.7
2:00 271.7 F 28.7
3:00 272.1 F 28.7
4:00 271.6 F 28.6

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 206.5 F 0.0
10:00 213.3 F 0.0
15:00 212.6 F 0.0
20:00 209.8 F 0.0
25:00 217.2 F 0.0
30:00 211.5 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:14
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

09/08/2023 10:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____