
Midmark M11 - v1.0.0
Total Cycles: 3811

Name: RITTER 1
Sterilizer ID: V2178198
09/25/2023 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 175.6 F 2.3
4:00 221.1 F 6.7
6:00 219.4 F 11.7
8:00 229.7 F 6.8
10:00 235.7 F 11.9
12:00 237.4 F 9.2
14:00 239.6 F 9.8
16:00 248.2 F 14.0
18:00 259.8 F 20.3
20:00 270.2 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.4
1:00 272.1 F 28.4
2:00 272.1 F 28.5
3:00 272.2 F 28.6
4:00 272.7 F 28.8

Min 271.4 F 28.3
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 201.8 F 0.0
10:00 204.5 F 0.0
15:00 206.7 F 0.0
20:00 209.6 F 0.0
25:00 200.8 F 0.0
Dry Cycle Aborted

FILLING: 1:49
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:37
DRYING: 28:50
TOTAL CYCLE: 00:59:38

09/25/2023 03:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____