
Midmark M11 - v1.0.0
Total Cycles: 1465

Name: RITTER 1
Sterilizer ID: V2178198
12/16/2021 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.9 F 0.0
2:00 138.0 F 0.0
4:00 166.0 F 0.1
6:00 184.7 F 2.5
8:00 210.5 F 6.6
10:00 225.8 F 10.9
12:00 227.0 F 6.0
14:00 238.5 F 10.8
16:00 233.0 F 7.4
18:00 245.1 F 12.5
20:00 244.6 F 12.4
22:00 257.2 F 19.0
24:00 268.4 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.5 F 28.9
2:00 271.8 F 28.2
3:00 271.3 F 28.3
4:00 272.5 F 28.9

Min 271.0 F 27.9
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 243.6 F 0.0
10:00 214.6 F 0.0
15:00 233.3 F 0.0
20:00 228.1 F 0.0
25:00 226.9 F 0.0
30:00 210.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:16
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:03:23
CYCLE COMPLETE

12/16/2021 10:59 AM
ACCEPTED: X
REJECTED:
Operator: U02