

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 000870  
15:39 14/01/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:22  
PRESSURIZING 3:27  
STERILIZING 4:39  
136.0 C 228kPa 4:39  
136.4 C 232kPa 5:09  
136.3 C 231kPa 5:39  
136.1 C 229kPa 6:09  
136.2 C 231kPa 6:39  
136.2 C 230kPa 7:09  
136.4 C 233kPa 7:39  
136.2 C 231kPa 8:09  
136.2 C 231kPa 8:09  
VENTING 8:09  
AIR DRYING 9:03  
CYCLE COMPLETE 24:03  
-----

Digital Signature #  
D8E320C14AE191D7  
-----