
Midmark M11 - v1.0.5
Total Cycles: 1070

Name: RITTER2
Sterilizer ID: V2313228
05/04/2022 10:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.1 F 0.0
2:00 151.3 F 2.4
4:00 168.0 F 6.0
6:00 191.1 F 10.8
8:00 212.9 F 4.4
10:00 222.8 F 9.9
12:00 231.6 F 8.2
14:00 241.0 F 12.7
16:00 241.1 F 11.5
18:00 254.1 F 18.2
20:00 265.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.5 F 30.1
2:00 272.3 F 29.6
3:00 271.2 F 29.2
4:00 272.0 F 30.1

Min 271.2 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.0 F 0.0
5:00 213.0 F 0.0
10:00 221.1 F 0.0
15:00 220.1 F 0.0
20:00 215.5 F 0.0
25:00 213.7 F 0.0
30:00 213.9 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:49
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

05/04/2022 11:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____