
Midmark M11 - v1.0.0
Total Cycles: 3995

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2023 11:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.9 F 0.0
2:00 167.9 F 2.7
4:00 184.1 F 6.6
6:00 204.7 F 11.5
8:00 221.3 F 5.0
10:00 230.3 F 9.6
12:00 236.7 F 9.3
14:00 241.8 F 11.6
16:00 239.1 F 9.8
18:00 252.1 F 16.1
20:00 263.2 F 22.5
22:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.3 F 28.6
2:00 272.8 F 29.0
3:00 271.5 F 28.6
4:00 272.3 F 28.7

Min 271.3 F 28.4
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 219.9 F 0.0
10:00 216.6 F 0.0
15:00 221.6 F 0.0
20:00 186.3 F 0.0
25:00 183.1 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:52
CYCLE COMPLETE

11/08/2023 12:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____