
Midmark M11 - v1.0.5
Total Cycles: 2854

Name: RITTER2
Sterilizer ID: V2313228
09/08/2023 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.2 F 0.0
2:00 136.3 F 0.1
4:00 161.7 F 2.9
6:00 187.4 F 6.8
8:00 201.5 F 11.5
10:00 218.6 F 6.3
12:00 231.0 F 11.2
14:00 231.6 F 8.2
16:00 241.7 F 11.9
18:00 244.2 F 12.8
20:00 255.9 F 19.0
22:00 266.4 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.6 F 29.8
2:00 271.8 F 29.1
3:00 271.4 F 29.3
4:00 272.6 F 29.7

Min 271.1 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.9 F 0.0
5:00 202.6 F 0.0
10:00 199.9 F 0.0
15:00 193.7 F 0.0
20:00 190.4 F 0.0
25:00 188.9 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

09/08/2023 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____