
Midmark M11 - v1.0.0
Total Cycles: 773

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2021 07:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 73.8 F 0.0
4:00 84.6 F 0.1
6:00 116.2 F 0.1
8:00 154.0 F 0.2
10:00 189.2 F 2.0
12:00 206.9 F 5.4
14:00 215.5 F 8.9
16:00 227.4 F 12.9
18:00 230.7 F 7.9
20:00 242.7 F 13.1
22:00 240.6 F 10.7
24:00 237.0 F 8.9
26:00 252.5 F 16.3
28:00 264.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.0 F 28.2
2:00 272.5 F 28.7
3:00 271.4 F 27.8
4:00 272.5 F 28.5

Min 271.3 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 171.9 F 0.0
5:00 195.5 F 0.0
10:00 197.8 F 0.0
15:00 196.4 F 0.0
20:00 192.1 F 0.0
25:00 192.8 F 0.0
Dry Cycle Aborted

FILLING: 0:49
HEATING: 29:48
STERILIZING: 4:00
VENTING: 2:52
DRYING: 27:54
TOTAL CYCLE: 01:05:22

04/07/2021 08:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____