
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002537
8:51 26/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:07
STERILIZING 4:24
136.0 C 220kPa 4:24
135.8 C 221kPa 4:54
136.2 C 223kPa 5:24
136.0 C 223kPa 5:54
136.4 C 225kPa 6:24
136.2 C 221kPa 6:54
136.4 C 224kPa 7:24
136.4 C 219kPa 7:54
136.4 C 219kPa 7:54
VENTING 7:54
AIR DRYING 8:45
CYCLE COMPLETE 10:50

Digital Signature #
A3372B5CF13329A2
