

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3078

Name: RITTER 1  
Sterilizer ID: V2178198  
04/10/2023 07:26 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.0 F 0.0  
2:00 77.7 F 0.0  
4:00 88.2 F 0.0  
6:00 111.2 F 0.1  
8:00 149.2 F 0.9  
10:00 178.2 F 3.2  
12:00 189.2 F 6.6  
14:00 204.2 F 10.2  
16:00 222.8 F 8.7  
18:00 227.2 F 7.5  
20:00 235.0 F 11.8  
22:00 234.7 F 8.4  
24:00 243.7 F 11.8  
26:00 245.0 F 12.5  
28:00 256.9 F 18.7  
30:00 267.5 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.9  
1:00 271.9 F 28.1  
2:00 272.3 F 28.4  
3:00 272.2 F 28.7  
4:00 271.4 F 27.9

Min 271.3 F 27.8  
Max 272.5 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
5:00 229.4 F 0.0  
10:00 207.6 F 0.0  
15:00 200.0 F 0.0  
20:00 207.1 F 0.0  
25:00 207.4 F 0.0  
30:00 229.2 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:21  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:09:28  
CYCLE COMPLETE

04/10/2023 08:35 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_