

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1140

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2022 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.1 F 0.0  
2:00 181.6 F 4.0  
4:00 198.5 F 9.1  
6:00 219.6 F 7.8  
8:00 232.3 F 10.6  
10:00 231.5 F 8.2  
12:00 240.5 F 11.1  
14:00 246.0 F 13.8  
16:00 258.4 F 20.5  
18:00 269.4 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.6 F 29.9  
2:00 271.8 F 29.1  
3:00 271.7 F 29.6  
4:00 272.9 F 29.8  
  
Min 271.3 F 29.0  
Max 273.0 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.5 F 0.0  
5:00 217.8 F 0.0  
10:00 228.1 F 0.0  
15:00 228.5 F 0.0  
20:00 225.4 F 0.0  
25:00 222.7 F 0.0  
30:00 218.5 F 0.0  
Drying Complete

FILLING: 2:03  
HEATING: 19:03  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:16  
CYCLE COMPLETE

05/31/2022 06:43 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02