

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4493

Name: RITTER 1  
Sterilizer ID: V2178198  
04/25/2024 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 88.6 F 0.0  
2:00 89.6 F 0.3  
4:00 96.2 F 2.0  
6:00 127.3 F 4.4  
8:00 172.3 F 7.5  
10:00 188.7 F 11.1  
12:00 211.9 F 3.3  
14:00 212.4 F 5.9  
16:00 224.0 F 9.6  
18:00 235.8 F 10.9  
20:00 237.0 F 10.2  
22:00 239.0 F 8.7  
24:00 246.9 F 13.7  
26:00 258.0 F 19.6  
28:00 267.9 F 25.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.1 F 28.7  
2:00 272.4 F 29.1  
3:00 271.4 F 28.4  
4:00 272.0 F 28.9

Min 271.2 F 28.3  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.1 F 0.0  
5:00 230.7 F 0.0  
10:00 207.0 F 0.0  
15:00 212.4 F 0.0  
20:00 207.8 F 0.0  
25:00 224.5 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:20  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:07:57  
CYCLE COMPLETE

04/25/2024 03:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_