

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3825

Name: RITTER 1  
Sterilizer ID: V2178198  
09/27/2023 08:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.4 F 0.0  
2:00 71.8 F 0.0  
4:00 91.2 F 0.0  
6:00 125.5 F 0.0  
8:00 155.1 F 0.2  
10:00 178.6 F 2.5  
12:00 207.5 F 5.3  
14:00 216.2 F 8.7  
16:00 225.9 F 12.4  
18:00 227.4 F 6.3  
20:00 237.3 F 10.8  
22:00 234.4 F 7.5  
24:00 246.1 F 13.2  
26:00 243.6 F 11.6  
28:00 255.9 F 18.0  
30:00 266.7 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.8  
1:00 272.3 F 28.2  
2:00 271.8 F 28.2  
3:00 272.1 F 28.0  
4:00 272.1 F 28.4

Min 271.3 F 27.6  
Max 272.6 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 228.9 F 0.0  
10:00 205.5 F 0.0  
15:00 201.6 F 0.0  
20:00 255.6 F 0.0  
25:00 235.6 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:32  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:09:34  
CYCLE COMPLETE

09/27/2023 09:42 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_