

-----  
Midmark M11 - v1.0.5  
Total Cycles: 368

Name: RITTER2  
Sterilizer ID: V2313228  
09/16/2021 06:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.6 F 0.0  
2:00 73.5 F 0.0  
4:00 77.4 F 0.0  
6:00 112.6 F 0.1  
8:00 152.5 F 0.3  
10:00 173.2 F 3.3  
12:00 190.2 F 6.6  
14:00 213.4 F 10.1  
16:00 219.7 F 10.8  
18:00 222.7 F 8.2  
20:00 234.6 F 12.5  
22:00 234.4 F 9.5  
24:00 238.5 F 10.1  
26:00 246.1 F 13.7  
28:00 256.8 F 19.7  
30:00 267.7 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.8  
1:00 272.4 F 29.6  
2:00 271.4 F 28.7  
3:00 272.0 F 29.6  
4:00 272.1 F 29.2

Min 271.0 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 218.0 F 0.0  
10:00 193.9 F 0.0  
15:00 190.5 F 0.0  
20:00 191.1 F 0.0  
25:00 192.1 F 0.0  
30:00 215.8 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:32  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:09:23  
CYCLE COMPLETE

09/16/2021 07:49 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02