
Midmark M11 - v1.0.5
Total Cycles: 245

Name: RITTER2
Sterilizer ID: V2313228
07/26/2021 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.1 F 0.0
2:00 170.6 F 3.3
4:00 188.0 F 7.9
6:00 209.0 F 13.2
8:00 219.5 F 9.0
10:00 232.0 F 10.3
12:00 238.0 F 12.6
14:00 240.2 F 11.6
16:00 252.9 F 18.1
18:00 264.2 F 24.8
20:00 271.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 30.0
1:00 271.7 F 30.3
2:00 272.0 F 30.5
3:00 271.9 F 30.4
4:00 272.2 F 30.4

Min 271.5 F 29.9
Max 272.6 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 239.8 F 0.0
10:00 230.6 F 0.0
15:00 226.1 F 0.0
20:00 221.8 F 0.0
25:00 228.0 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:45
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

07/26/2021 11:25 AM
ACCEPTED: X
REJECTED:
Operator: U01